



Deadtime Analysis Progress


9/26/2017

Salina Ali, Mongi Dlamini

Trigger Setup



DVCS Calorimeter



Borrowed from <https://hallaweb.jlab.org/wiki/index.php/Trigger>

Deadtime Computations

3

- Looking at scaler rates: live and raw

$$\text{Raw rate} = \text{Live rate} \cdot \frac{1}{1 - \text{Dedtime}}$$

- Livetime and dedtime:

$$\text{Livetime} = \frac{\text{Live Scaler Rate}}{\text{Raw Scaler Rate}}$$

$$\text{Dedtime} = 1 - \text{Livetime}$$

Scaler check: Master OR, ARS Stop

- Scalers are **not** “randomly counting”:
 - Checking ARS Stop and Master OR live- these are consistently “double” counting..which means that they are in agreement.
 - S2M && Cer Live events are NOT validated, so they should not count the way ARS Valid does.

```




File name : left_dvcs_13418.root
root [2] T->Scan("gated_accum_dvcs_scaler_24")
*****
*   Row   * gated_acc *
*****
*     0 *         1 *
*     1 *         2 *
*     2 *         4 *
*     3 *         5 *
*     4 *         7 *
*     5 *         8 *
*     6 *         9 *
*     7 *        11 *
*     8 *        12 *
*     9 *        13 *
*    10 *        14 *
*    11 *        15 *
*    12 *        16 *
*    13 *        17 *
*    14 *        18 *
*    15 *        19 *
*    16 *        20 *
*    17 *        21 *
*    18 *        22 *
*    19 *        24 *
*    20 *        25 *
*    21 *        26 *
*    22 *        27 *
*    23 *        28 *
*    24 *        30 *
Type <CR> to continue or q to quit ==>
*    25 *        31 *
*    26 *        32 *
*    27 *        33 *
*    28 *        34 *
*    29 *        35 *
*    30 *        37 *
*    31 *        39 *
*    32 *        40 *
*    33 *        41 *
*    34 *        42 *
*    35 *        43 *

```

```



root [3] T->Scan("gated_accum_dvcs_scaler_ARSStop")
*****
*   Row   * gated_acc *
*****
*     0 *         1 *
*     1 *         2 *
*     2 *         4 *
*     3 *         5 *
*     4 *         7 *
*     5 *         8 *
*     6 *         9 *
*     7 *        11 *
*     8 *        12 *
*     9 *        13 *
*    10 *        14 *
*    11 *        15 *
*    12 *        16 *
*    13 *        17 *
*    14 *        18 *
*    15 *        19 *
*    16 *        20 *
*    17 *        21 *
*    18 *        22 *
*    19 *        24 *
*    20 *        25 *
*    21 *        26 *
*    22 *        27 *
*    23 *        28 *
*    24 *        30 *
Type <CR> to continue or q to quit ==>
*    25 *        31 *
*    26 *        32 *
*    27 *        33 *
*    28 *        34 *
*    29 *        35 *
*    30 *        37 *
*    31 *        39 *
*    32 *        40 *
*    33 *        41 *
*    34 *        42 *
*    35 *        43 *
*    36 *        44 *
*    37 *        45 *
*    38 *        46 *
*    39 *        47 *

```



ARS stop, Master OR
Scalers are consistent.

S2m&&Cer, ARS Valid



- Signal after s2m&&cer is not validated by ARS, which is why scalers are different.

```

root [5] T->Scan("cptARSValid")
*****
*      Row      * cptARSVal *
*****
*      0 *      0 *
*      1 *      1 *
*      2 *      2 *
*      3 *      3 *
*      4 *      4 *
*      5 *      5 *
*      6 *      6 *
*      7 *      7 *
*      8 *      8 *
*      9 *      9 *
*     10 *     10 *
*     11 *     11 *
*     12 *     12 *
*     13 *     13 *
*     14 *     14 *
*     15 *     15 *
*     16 *     16 *
*     17 *     17 *
*     18 *     18 *
*     19 *     19 *
*     20 *     20 *
*     21 *     21 *
*     22 *     22 *
*     23 *     23 *
*     24 *     24 *
Type <CR> to continue or q to quit ==>
*     25 *     25 *
*     26 *     26 *
*     27 *     27 *
*     28 *     28 *
*     29 *     29 *
*     30 *     30 *
*     31 *     31 *
*     32 *     32 *
*     33 *     33 *
*     34 *     34 *
*     35 *     35 *
*     36 *     36 *
*     37 *     37 *
*     38 *     38 *
*     39 *     39 *

```

```

(Long64 [1]30
root [4] T->Scan("cptARSSto")
*****
*      Row      * cptARSSto *
*****
*      0 *      1 *
*      1 *      2 *
*      2 *      4 *
*      3 *      5 *
*      4 *      7 *
*      5 *      8 *
*      6 *      9 *
*      7 *     11 *
*      8 *     12 *
*      9 *     13 *
*     10 *     14 *
*     11 *     15 *
*     12 *     16 *
*     13 *     17 *
*     14 *     18 *
*     15 *     19 *
*     16 *     20 *
*     17 *     21 *
*     18 *     22 *
*     19 *     24 *
*     20 *     25 *
*     21 *     26 *
*     22 *     27 *
*     23 *     28 *
*     24 *     30 *
Type <CR> to continue or q to quit ==>
*     25 *     31 *
*     26 *     32 *
*     27 *     33 *
*     28 *     34 *
*     29 *     35 *
*     30 *     37 *
*     31 *     39 *
*     32 *     40 *
*     33 *     41 *
*     34 *     42 *
*     35 *     43 *
*     36 *     44 *
*     37 *     45 *

```

```

(Long64 [1]30
root [3] T->Scan("gated_accum_dvcs_scaler_ARSSto")
*****
*      Row      * gated_acc *
*****
*      0 *      1 *
*      1 *      2 *
*      2 *      4 *
*      3 *      5 *
*      4 *      7 *
*      5 *      8 *
*      6 *      9 *
*      7 *     11 *
*      8 *     12 *
*      9 *     13 *
*     10 *     14 *
*     11 *     15 *
*     12 *     16 *
*     13 *     17 *
*     14 *     18 *
*     15 *     19 *
*     16 *     20 *
*     17 *     21 *
*     18 *     22 *
*     19 *     24 *
*     20 *     25 *
*     21 *     26 *
*     22 *     27 *
*     23 *     28 *
*     24 *     30 *
Type <CR> to continue or q to quit ==>
*     25 *     31 *
*     26 *     32 *
*     27 *     33 *
*     28 *     34 *
*     29 *     35 *
*     30 *     37 *
*     31 *     39 *
*     32 *     40 *
*     33 *     41 *
*     34 *     42 *
*     35 *     43 *
*     36 *     44 *
*     37 *     45 *
*     38 *     46 *
*     39 *     47 *

```

Normalized DVCS and DIS rates

Normalized Rates (Hz/uA)											
I(uA)	S2M &Cer LT	No cuts	Trk	Trk&TD C	Trk&TD C&Cer	Trk&TD C&Cer& DIS	Trk&TDC& Cer&DIS/S 2M&CER LT	Trk&TDC&Cer&DVC er&DVCS/S 2M&CER LT	Trk&TDC&Ce r&DVCS*corr/ S2M&CER LT		
10.61	0.985	9.27	5.783	5.719	5.138	3.365	3.422	5.134	5.212		4.125
15.32	0.976	10.26	6.192	6.117	5.484	3.356	3.450	5.480	5.615		4.154
20.53	0.965	11.26	6.459	6.391	5.733	3.321	3.449	5.728	5.936		3.886


- Rates given in Hz/uA, with the following cuts:
 - Trk: tracking cut, given by “L.tr.n” >0
 - TDC: Time-to-Digital Converter, given by tdc_val[27]-tdc_val[7]/10<-24
 - CER: Cerenkov cut, given by “L.cer >500”
 - DIS: given by “triggerPatternWord&0x00080”
 - DVCS: given by “triggerPatternWord&0x00100”

Current dependence goes away..




Mult by correction factor

Waveform Analysis → Clustering


- Performed clustering to account for accidentals: requires waveform analysis, and then applying calorimeter data to get true and accidental coincidences
 - t_{moy} corresponds to time spectrum of blocks $E_\gamma > 1\text{GeV}$, in which **main coincidence peak** tells us photon + electron + background.
 - **Accidental region** is where the accidentals are in 4ns blocks.
- Need to subtract accidental region from coincidence window to correct for accidentals.



Coincidence time spectrum

tmoy: ntr,cer,tdcval, trigpatW&0x00100==256 cuts at 10 μ Atmoy: ntr,cer,tdcval, trigpatW&0x00100==256 cuts at 15 μ Atmoy: ntr,cer,tdcval, trigpatW&0x00100==256 cuts at 20 μ A

- Low statistics; applied cerenkov>500, trk<0, TDC , trigpatW cuts.

tmoy: ntr,cer,tdcval, trigpatW cuts at 10 μA tmoy: ntr,cer,tdcval, trigpatW cuts at 10 μA 